

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
 Data File : VD074320.D
 Acq On : 15 Sep 2022 20:47
 Operator : VA/SY
 Sample : N4582-01RE
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CRUSHED-MASONRY-PILE-7RE

Quant Time: Sep 16 05:56:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	128872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	227263	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	198642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	87668	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	60487	46.037	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.080%
35) Dibromofluoromethane	7.810	113	22297	12.166	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	24.340%#
50) Toluene-d8	10.269	98	286581	51.442	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.880%
62) 4-Bromofluorobenzene	12.569	95	81074	40.331	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.660%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.798	84	23875	9.542	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
Data File : VD074320.D
Acq On : 15 Sep 2022 20:47
Operator : VA/SY
Sample : N4582-01RE
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CRUSHED-MASONRY-PILE-7RE

Quant Time: Sep 16 05:56:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

