

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091620\
 Data File : VD066718.D
 Acq On : 16 Sep 2020 14:36
 Operator : VA/SY
 Sample : L4007-01
 Misc : 3.02G/5.00ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-11929

Quant Time: Sep 17 01:08:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 16:35:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	377130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	708313	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	669369	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	291150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	244914	60.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.74%	
35) Dibromofluoromethane	7.91	113	101402	23.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	46.70%	
50) Toluene-d8	10.34	98	865471	55.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.88%	
62) 4-Bromofluorobenzene	12.63	95	295250	55.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.56%	
Target Compounds						
8) Diethyl Ether	3.70	74	2411	1.298	ug/l	91
16) Acetone	4.16	43	21755	26.550	ug/l	96
20) Methylene Chloride	4.97	84	70934	10.252	ug/l	96
52) Toluene	10.40	92	67132	5.319	ug/l	100
95) Naphthalene	15.41	128	10859	1.248	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091620\
Data File : VD066718.D
Acq On : 16 Sep 2020 14:36
Operator : VA/SY
Sample : L4007-01
Misc : 3.02G/5.00ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-11929

Quant Time: Sep 17 01:08:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 16:35:37 2020
Response via : Initial Calibration

