

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091620\
 Data File : VD066719.D
 Acq On : 16 Sep 2020 15:04
 Operator : VA/SY
 Sample : L4020-01
 Misc : 4.87G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-362

Quant Time: Sep 17 01:11:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 16:35:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	349852	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	646536	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	611405	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	271012	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	229003	61.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.70%	
35) Dibromofluoromethane	7.91	113	225737	56.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.90%	
50) Toluene-d8	10.34	98	789835	55.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.86%	
62) 4-Bromofluorobenzene	12.63	95	267291	54.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.66%	
Target Compounds						
8) Diethyl Ether	3.71	74	3144	1.824	ug/l	92
16) Acetone	4.16	43	112802	148.399	ug/l	90
20) Methylene Chloride	4.96	84	79488	14.179	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091620\
Data File : VD066719.D
Acq On : 16 Sep 2020 15:04
Operator : VA/SY
Sample : L4020-01
Misc : 4.87G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-362

Quant Time: Sep 17 01:11:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 16:35:37 2020
Response via : Initial Calibration

