


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD091620\
 Data File : VD066712.D
 Acq On : 16 Sep 2020 10:41
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Title : SW846 8260
 Last Update : Tue Sep 15 16:35:37 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	45082	PASS
75	95	30	60	56.9	117085	PASS
95	95	100	100	100.0	205674	PASS
96	95	5	9	7.0	14408	PASS
173	174	0.00	2	0.8	1243	PASS
174	95	50	100	76.2	156629	PASS
175	174	5	9	7.6	11859	PASS
176	174	95	101	95.2	149056	PASS
177	176	5	9	7.2	10746	PASS