

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
 Data File : VD074330.D
 Acq On : 16 Sep 2022 14:53
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 17 04:22:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	105340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	181208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	159245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	73220	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	53835	52.469	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.940%
35) Dibromofluoromethane	7.805	113	54292	45.492	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.980%
50) Toluene-d8	10.269	98	230648	56.010	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.020%
62) 4-Bromofluorobenzene	12.569	95	66357	47.231	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
Data File : VD074330.D
Acq On : 16 Sep 2022 14:53
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Sep 17 04:22:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

