

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
 Data File : VD074332.D
 Acq On : 16 Sep 2022 16:19
 Operator : VA/SY
 Sample : VD0916SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0916SBL01

Quant Time: Sep 17 04:23:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	101592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	172536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	153634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	69380	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	52582	53.139	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.280%
35) Dibromofluoromethane	7.804	113	52982	46.625	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.260%
50) Toluene-d8	10.269	98	217596	55.496	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.000%
62) 4-Bromofluorobenzene	12.569	95	62120	46.329	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
Data File : VD074332.D
Acq On : 16 Sep 2022 16:19
Operator : VA/SY
Sample : VD0916SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0916SBL01

Quant Time: Sep 17 04:23:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

