

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
 Data File : VD074330.D
 Acq On : 16 Sep 2022 14:53
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 17 04:22:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	105340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	181208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	159245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	73220	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	53835	52.469	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.940%
35) Dibromofluoromethane	7.805	113	54292	45.492	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.980%
50) Toluene-d8	10.269	98	230648	56.010	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.020%
62) 4-Bromofluorobenzene	12.569	95	66357	47.231	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.460%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
Data File : VD074330.D
Acq On : 16 Sep 2022 14:53
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Sep 17 04:22:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1109
Delta R.T. 0.000 min
Lab File: VD074330.D
Acq: 16 Sep 2022 14:53




Instrument :
MSVOA_D
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 105340
Ion Ratio Lower Upper
168 100
99 57.7 43.0 64.6



#33
1,2-Dichloroethane-d4
Concen: 52.469 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.006 min
Lab File: VD074330.D
Acq: 16 Sep 2022 14:53

Tgt Ion: 65 Resp: 53835
Ion Ratio Lower Upper
65 100
67 54.6 0.0 110.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD074330.D

Acq: 16 Sep 2022 14:53

Instrument :

MSVOA_D

ClientSampleId :

VIBLK





Tgt Ion:114 Resp: 181208

Ion Ratio Lower Upper

114 100

63 19.3 0.0 37.6

88 15.7 0.0 28.4



#35

Dibromofluoromethane

Concen: 45.492 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. 0.001 min

Lab File: VD074330.D

Acq: 16 Sep 2022 14:53




Tgt Ion:113 Resp: 54292


Ion Ratio Lower Upper

113 100

111 105.9 83.0 124.6

192 19.2 16.2 24.2









#50
Toluene-d8
Concen: 56.010 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074330.D
Acq: 16 Sep 2022 14:53




Instrument :
MSVOA_D
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 230648
Ion Ratio Lower Upper
98 100
100 64.0 51.8 77.8



#62
4-Bromofluorobenzene
Concen: 47.231 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VD074330.D
Acq: 16 Sep 2022 14:53

Tgt Ion: 95 Resp: 66357
Ion Ratio Lower Upper
95 100
174 88.3 0.0 179.2
176 84.6 0.0 171.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.575 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074330.D
Acq: 16 Sep 2022 14:53

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Tgt Ion:117 Resp: 159245
Ion Ratio Lower Upper
117 100
82 54.5 44.2 66.2
119 30.9 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.510 min Scan# 2027
Delta R.T. 0.000 min
Lab File: VD074330.D
Acq: 16 Sep 2022 14:53

Tgt Ion:152 Resp: 73220
Ion Ratio Lower Upper
152 100
115 55.1 28.4 85.2
150 157.1 0.0 351.0

