

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
 Data File : VD074338.D
 Acq On : 16 Sep 2022 19:07
 Operator : VA/SY
 Sample : N4613-01
 Misc : 5.39G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-6-(10-18)

Quant Time: Sep 17 04:24:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	118833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	202974	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	176342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	75488	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	51388	44.397	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.800%
35) Dibromofluoromethane	7.810	113	54027	40.415	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.840%
50) Toluene-d8	10.269	98	219637	47.617	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.240%
62) 4-Bromofluorobenzene	12.569	95	59133	36.257	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091622\
Data File : VD074338.D
Acq On : 16 Sep 2022 19:07
Operator : VA/SY
Sample : N4613-01
Misc : 5.39G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-6-(10-18)

Quant Time: Sep 17 04:24:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074338.D
Acq: 16 Sep 2022 19:07




Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-6-(10-18)


Tgt Ion:168 Resp: 118833
Ion Ratio Lower Upper
168 100
99 56.9 43.0 64.6



#33
1,2-Dichloroethane-d4
Concen: 44.397 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.006 min
Lab File: VD074338.D
Acq: 16 Sep 2022 19:07

Tgt Ion: 65 Resp: 51388
Ion Ratio Lower Upper
65 100
67 54.0 0.0 110.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 11

Delta R.T. -0.006 min

Lab File: VD074338.D

Acq: 16 Sep 2022 19:07

Instrument :

MSVOA_D

ClientSampleId :

SASP2-T-6-(10-18)





Tgt Ion:114 Resp: 202974

Ion Ratio Lower Upper

114 100

63 21.6 0.0 37.6

88 16.2 0.0 28.4



#35

Dibromofluoromethane

Concen: 40.415 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VD074338.D

Acq: 16 Sep 2022 19:07




Tgt Ion:113 Resp: 54027


Ion Ratio Lower Upper

113 100

111 101.7 83.0 124.6

192 17.8 16.2 24.2









#50
Toluene-d8
Concen: 47.617 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074338.D
Acq: 16 Sep 2022 19:07




Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-6-(10-18)


Tgt Ion: 98 Resp: 219637
Ion Ratio Lower Upper
98 100
100 61.8 51.8 77.8



#62
4-Bromofluorobenzene
Concen: 36.257 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD074338.D
Acq: 16 Sep 2022 19:07

Tgt Ion: 95 Resp: 59133
Ion Ratio Lower Upper
95 100
174 88.5 0.0 179.2
176 86.5 0.0 171.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VD074338.D
Acq: 16 Sep 2022 19:07

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-6-(10-18)

Tgt Ion:117 Resp: 176342
Ion Ratio Lower Upper
117 100
82 52.4 44.2 66.2
119 32.2 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.510 min Scan# 2027
Delta R.T. 0.000 min
Lab File: VD074338.D
Acq: 16 Sep 2022 19:07

Tgt Ion:152 Resp: 75488
Ion Ratio Lower Upper
152 100
115 56.8 28.4 85.2
150 157.0 0.0 351.0

