

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091720\
 Data File : VD066743.D
 Acq On : 17 Sep 2020 14:37
 Operator : VA/SY
 Sample : L4033-23
 Misc : 8.50G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 78-80

Quant Time: Sep 18 01:15:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 16:35:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	395761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	701440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	676754	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	295342	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.33	65	230678	54.63	ug/l	0.00
Spiked Amount			Recovery	=	109.26%	
35) Dibromofluoromethane	7.91	113	236057	54.89	ug/l	0.00
Spiked Amount			Recovery	=	109.78%	
50) Toluene-d8	10.34	98	870918	56.84	ug/l	0.00
Spiked Amount			Recovery	=	113.68%	
62) 4-Bromofluorobenzene	12.63	95	304073	57.49	ug/l	0.00
Spiked Amount			Recovery	=	114.98%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.99	85	306	Below Cal	#	43
11) Tert butyl alcohol	5.21	59	3458	Below Cal	#	89
16) Acetone	4.17	43	22307	25.942	ug/l	95
17) Carbon Disulfide	4.40	76	49615	3.987	ug/l	95
18) Methyl Acetate	4.72	43	326	Below Cal		94
20) Methylene Chloride	4.96	84	97842	16.189	ug/l #	86
25) 2-Butanone	7.23	43	7798	6.964	ug/l #	74
31) Cyclohexane	7.98	56	15286	2.049	ug/l #	40
36) 1,1-Dichloropropene	7.97	75	35920	4.981	ug/l #	52
37) Ethyl Acetate	7.23	43	7798	2.856	ug/l #	70
38) Carbon Tetrachloride	7.98	117	44517	5.813	ug/l #	17
39) Methylcyclohexane	9.35	83	9934	1.204	ug/l #	83
40) Benzene	8.35	78	125161	6.415	ug/l	95
51) 4-Methyl-2-Pentanone	10.23	43	4783	1.808	ug/l	93
52) Toluene	10.40	92	172079	13.767	ug/l	98
67) Ethyl Benzene	11.74	91	76273	3.005	ug/l	94
68) m/p-Xylenes	11.85	106	98246	10.242	ug/l	96
69) o-Xylene	12.18	106	45462	5.363	ug/l	96
70) Styrene	12.20	104	16398	1.099	ug/l #	54
74) N-amyl acetate	12.23	43	23877	5.134	ug/l #	48
76) 1,2,3-Trichloropropane	12.63	75	169040	66.867	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.96	105	26136	1.402	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.63	75	169040	144.310	ug/l #	8
84) 1,2,4-Trimethylbenzene	13.27	105	48041	2.579	ug/l	98
95) Naphthalene	15.41	128	150566	17.056	ug/l #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091720\
Data File : VD066743.D
Acq On : 17 Sep 2020 14:37
Operator : VA/SY
Sample : L4033-23
Misc : 8.50G/5.00ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
78-80

Quant Time: Sep 18 01:15:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 16:35:37 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampleId :


78-80

Tot Ion:168 Resp: 395761

Ion Ratio Lower Upper

168 100

99 61.1 54.4 81.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98



150000

100000

50000

0

Time-->



#2

Dichlorodifluoromethane

Concen: Below Cal

RT: 1.99 min Scan# 11

Delta R.T. -0.01 min

Lab File: VD066743.D


Acq: 17 Sep 2020 14:37

Tgt Ion: 85 Resp: 306

Ion Ratio Lower Upper

85 100

87 0.0 15.9 47.6#



Abundance Ion 84.90 (84.60 to 85.60): VDC

Ion 86.90 (86.60 to 87.60): VDC

1.99

300

250

200


150


100

50

0

Time-->









#11

Tert butyl alcohol
 Concen: Below Cal
 RT: 5.21 min Scan# 560
 Delta R.T. -0.01 min
 Lab File: VD066743.D
 Acq: 17 Sep 2020 14:37

Instrument :
 MSVOA_D
 ClientSampleId :
 78-80




Tot Ion: 59 Resp: 3458
 Ion Ratio Lower Upper
 59 100
 57 4.4 6.7 10.1#




#16

Acetone
 Concen: 25.942 ug/l
 RT: 4.17 min Scan# 382
 Delta R.T. 0.01 min
 Lab File: VD066743.D
 Acq: 17 Sep 2020 14:37

Tgt Ion: 43 Resp: 22307
 Ion Ratio Lower Upper
 43 100
 58 25.0 21.9 32.9









#17
Carbon Disulfide
Concen: 3.987 ug/l
RT: 4.40 min Scan# 422
Delta R.T. -0.01 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37




Instrument :
MSVOA_D
ClientSampled :
78-80


Tot Ion: 76 Resp: 49615
Ion Ratio Lower Upper
76 100
78 8.2 8.0 12.0



#18
Methyl Acetate
Concen: Below Cal
RT: 4.72 min Scan# 476
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Tgt Ion: 43 Resp: 326
Ion Ratio Lower Upper
43 100
74 20.6 18.6 28.0









#20
Methylene Chloride
Concen: 16.189 ug/l
RT: 4.96 min Scan# 517
Delta R.T. -0.01 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Instrument :
MSVOA_D
ClientSampleId :
78-80

Tot Ion: 84 Resp: 97842
Ion Ratio Lower Upper
84 100
49 109.6 105.0 157.6
51 32.0 32.2 48.4#
86 59.4 50.4 75.6




Abundance Ion 83.90 (83.60 to 84.60): VDC
Ion 48.90 (48.60 to 49.60): VDC
Ion 50.90 (50.60 to 51.60): VDC
Ion 85.90 (85.60 to 86.60): VDC






#25
2-Butanone
Concen: 6.964 ug/l
RT: 7.23 min Scan# 902
Delta R.T. 0.02 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Tgt Ion: 43 Resp: 7798
Ion Ratio Lower Upper
43 100
72 39.5 21.0 31.4#



Abundance Ion 43.00 (42.70 to 43.70): VDC
Ion 72.00 (71.70 to 72.70): VDC









#31
Cyclohexane
Concen: 2.049 ug/l
RT: 7.98 min Scan# 1030
Delta R.T. -0.01 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Instrument :
MSVOA_D
ClientSampled :
78-80

Tot Ion: 56 Resp: 15286
Ion Ratio Lower Upper
56 100
69 111.5 27.7 41.5#
84 114.1 68.5 102.7#




Abundance Ion 56.00 (55.70 to 56.70): VDC
Ion 69.00 (68.70 to 69.70): VDC
Ion 84.00 (83.70 to 84.70): VDC






#33
1,2-Dichloroethane-d4
Concen: 54.630 ug/l
RT: 8.33 min Scan# 1090
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Tgt Ion: 65 Resp: 230678
Ion Ratio Lower Upper
65 100
67 51.5 0.0 102.2



Abundance Ion 64.90 (64.60 to 65.60): VDC
Ion 66.90 (66.60 to 67.60): VDC









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.86 min Scan# 1180
 Delta R.T. 0.00 min
 Lab File: VD066743.D
 Acq: 17 Sep 2020 14:37

Instrument :
 MSVOA_D
 ClientSampleId :
 78-80

Tot Ion:114 Resp: 701440
 Ion Ratio Lower Upper
 114 100
 63 21.3 0.0 37.6
 88 16.4 0.0 32.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 54.891 ug/l
 RT: 7.91 min Scan# 1018
 Delta R.T. 0.00 min
 Lab File: VD066743.D
 Acq: 17 Sep 2020 14:37

Tgt Ion:113 Resp: 236057
 Ion Ratio Lower Upper
 113 100
 111 102.8 83.2 124.8
 192 17.3 14.8 22.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.981 ug/l

RT: 7.97 min Scan# 1029

Delta R.T. -0.14 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampled :

78-80





Tot Ion: 75 Resp: 35920

Ion Ratio Lower Upper

75 100

110 11.2 17.1 51.2#

77 0.0 25.0 37.6#



#37

Ethyl Acetate

Concen: 2.856 ug/l

RT: 7.23 min Scan# 902

Delta R.T. -0.06 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37




Tgt Ion: 43 Resp: 7798


Ion Ratio Lower Upper

43 100

61 0.0 10.4 15.6#

70 0.0 7.5 11.3#





#38

Carbon Tetrachloride

Concen: 5.813 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. -0.12 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampled :

78-80


Tot Ion:117 Resp: 44517

Ion Ratio Lower Upper

117 100

119 5.7 75.8 113.6#




121 0.0 26.2 39.4#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#39

Methylcyclohexane

Concen: 1.204 ug/l

RT: 9.35 min Scan# 1263

Delta R.T. -0.01 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37


Tgt Ion: 83 Resp: 9934

Ion Ratio Lower Upper

83 100

55 60.8 61.5 92.3#



98 39.8 39.3 58.9




Abundance Ion 83.00 (82.70 to 83.70): VDC

Ion 55.00 (54.70 to 55.70): VDC

Ion 98.00 (97.70 to 98.70): VDC





#40

Benzene

Concen: 6.415 ug/l

RT: 8.35 min Scan# 1093

Delta R.T. 0.00 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampleId :





78-80

Tot Ion: 78 Resp: 125161

Ion Ratio Lower Upper

78 100

77 26.7 19.3 28.9



#50

Toluene-d8

Concen: 56.842 ug/l

RT: 10.34 min Scan# 1431

Delta R.T. 0.00 min

Lab File: VD066743.D




Acq: 17 Sep 2020 14:37


Tgt Ion: 98 Resp: 870918

Ion Ratio Lower Upper

98 100

100 65.3 52.7 79.1









#51
 4-Methyl-2-Pentanone
 Concen: 1.808 ug/l
 RT: 10.23 min Scan# 1412
 Delta R.T. 0.00 min
 Lab File: VD066743.D
 Acq: 17 Sep 2020 14:37




Instrument :
 MSVOA_D
 ClientSampled :
 78-80


Tot Ion: 43 Resp: 4783
 Ion Ratio Lower Upper
 43 100
 58 33.9 30.4 45.6



#52
 Toluene
 Concen: 13.767 ug/l
 RT: 10.40 min Scan# 1442
 Delta R.T. 0.00 min
 Lab File: VD066743.D
 Acq: 17 Sep 2020 14:37

Tgt Ion: 92 Resp: 172079
 Ion Ratio Lower Upper
 92 100
 91 172.0 135.8 203.8









#62
4-Bromofluorobenzene
Concen: 57.493 ug/l
RT: 12.63 min Scan# 1821
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37




Instrument :
MSVOA_D
ClientSampleId :
78-80


Tot Ion: 95 Resp: 304073
Ion Ratio Lower Upper
95 100
174 75.4 0.0 156.6
176 74.2 0.0 150.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.64 min Scan# 1653
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Tgt Ion: 117 Resp: 676754
Ion Ratio Lower Upper
117 100
82 57.6 45.4 68.0
119 32.6 25.8 38.6









#67
Ethyl Benzene
Concen: 3.005 ug/l
RT: 11.74 min Scan# 1670
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37




Instrument :
MSVOA_D
ClientSampleId :
78-80


Tot Ion: 91 Resp: 76273
Ion Ratio Lower Upper
91 100
106 34.5 25.1 37.7



#68
m/p-Xylenes
Concen: 10.242 ug/l
RT: 11.85 min Scan# 1688
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Tgt Ion: 106 Resp: 98246
Ion Ratio Lower Upper
106 100
91 199.3 164.8 247.2









#69
o-Xylene
Concen: 5.363 ug/l
RT: 12.18 min Scan# 1745
Delta R.T. 0.01 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Instrument :
MSVOA_D
ClientSampleId :
78-80

Tot Ion:106 Resp: 45462
Ion Ratio Lower Upper
106 100
91 225.6 109.4 328.2




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): V






#70
Styrene
Concen: 1.099 ug/l
RT: 12.20 min Scan# 1747
Delta R.T. 0.00 min
Lab File: VD066743.D
Acq: 17 Sep 2020 14:37

Tgt Ion:104 Resp: 16398
Ion Ratio Lower Upper
104 100
78 87.9 40.2 60.4#
103 79.7 42.5 63.7#



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): V
Ion 103.00 (102.70 to 103.70): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.58 min Scan# 1982

Delta R.T. 0.00 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampleId :

78-80


Tot Ion:152 Resp: 295342

Ion Ratio Lower Upper

152 100

115 62.0 30.5 91.5




150 160.6 0.0 344.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 5.134 ug/l

RT: 12.23 min Scan# 1752

Delta R.T. -0.06 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Tgt Ion: 43 Resp: 23877


Ion Ratio Lower Upper

43 100

70 1.3 33.0 49.4#

55 3.0 20.6 30.8#

61 4.5 19.1 28.7#






Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 70.00 (69.70 to 70.70): VDC

Ion 55.00 (54.70 to 55.70): VDC

Ion 61.00 (60.70 to 61.70): VDC





#76

1,2,3-Trichloropropane

Concen: 66.867 ug/l

RT: 12.63 min Scan# 1821

Delta R.T. -0.15 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampleId :


78-80

Tot Ion: 75 Resp: 169040

Ion Ratio Lower Upper

75 100

77 1.5 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VD066698.D (-1845) (-)

Ion 77.00 (76.70 to 77.70): VD066698.D (-1845) (-)

120000

100000

80000

60000



40000

20000

0

12.50 12.60 12.70

Time-->



#80

1,3,5-Trimethylbenzene

Concen: 1.402 ug/l

RT: 12.96 min Scan# 1877

Delta R.T. 0.00 min

Lab File: VD066743.D


Acq: 17 Sep 2020 14:37

Tgt Ion: 105 Resp: 26136

Ion Ratio Lower Upper

105 100

120 47.1 23.9 71.7



Abundance Ion 105.00 (104.70 to 105.70): VD066698.D (-1869) (-)

Ion 120.00 (119.70 to 120.70): VD066698.D (-1869) (-)

20000

15000


10000


5000

0

12.90 12.95 13.00

Time-->





#81

trans-1,4-Dichloro-2-butene

Concen: 144.310 ug/l

RT: 12.63 min Scan# 1821

Delta R.T. 0.10 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampleId :

78-80


Tot Ion: 75 Resp: 169040

Ion Ratio Lower Upper

75 100

53 0.2 82.2 123.4#




89 0.2 39.6 59.4#



Abundance Ion 74.90 (74.60 to 75.60): VDC

Ion 53.00 (52.70 to 53.70): VDC

Ion 88.90 (88.60 to 89.60): VDC



#84

1,2,4-Trimethylbenzene

Concen: 2.579 ug/l

RT: 13.27 min Scan# 1930

Delta R.T. 0.00 min

Lab File: VD066743.D


Acq: 17 Sep 2020 14:37

Tgt Ion: 105 Resp: 48041

Ion Ratio Lower Upper



105 100


120 46.2 22.4 67.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#95

Naphthalene

Concen: 17.056 ug/l

RT: 15.41 min Scan# 2293

Delta R.T. 0.00 min

Lab File: VD066743.D

Acq: 17 Sep 2020 14:37

Instrument :

MSVOA_D

ClientSampleId :

78-80

Tot Ion:128 Resp: 150566

Ion Ratio Lower Upper

128 100

127 15.4 10.5 15.7

129 13.6 8.6 12.8#

