


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD091720\
 Data File : VD066733.D
 Acq On : 17 Sep 2020 09:24
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Title : SW846 8260
 Last Update : Tue Sep 15 16:35:37 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	35570	PASS
75	95	30	60	54.2	92992	PASS
95	95	100	100	100.0	171498	PASS
96	95	5	9	6.8	11727	PASS
173	174	0.00	2	1.8	2422	PASS
174	95	50	100	77.5	132893	PASS
175	174	5	9	7.4	9830	PASS
176	174	95	101	96.8	128586	PASS
177	176	5	9	6.8	8798	PASS