

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091823\
 Data File : VD077126.D
 Acq On : 18 Sep 2023 13:08
 Operator : JC/SY
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 19 04:54:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	174061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	386462	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	407471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.521	152	194729	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	94362	53.717	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.440%
35) Dibromofluoromethane	7.810	113	130041	48.444	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.880%
50) Toluene-d8	10.269	98	465955	48.767	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.540%
62) 4-Bromofluorobenzene	12.574	95	160924	54.340	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091823\
Data File : VD077126.D
Acq On : 18 Sep 2023 13:08
Operator : JC/SY
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Sep 19 04:54:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

