

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091823\
 Data File : VD077132.D
 Acq On : 18 Sep 2023 16:21
 Operator : JC/SY
 Sample : 04436-11
 Misc : 5.26G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B102-11(3.5-4)

Quant Time: Sep 19 04:57:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	140842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	318639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	344181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	159112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	77152	54.279	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.560%	
35) Dibromofluoromethane	7.810	113	112012	50.610	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.220%	
50) Toluene-d8	10.269	98	391802	49.735	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.460%	
62) 4-Bromofluorobenzene	12.575	95	128070	52.451	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 104.900%	
Target Compounds						
					Qvalue	
16) Acetone	4.028	43	7854	24.176	ug/l	99
20) Methylene Chloride	4.805	84	35970	12.360	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091823\
Data File : VD077132.D
Acq On : 18 Sep 2023 16:21
Operator : JC/SY
Sample : 04436-11
Misc : 5.26G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B102-11(3.5-4)

Quant Time: Sep 19 04:57:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

