

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091823\
 Data File : VD077138.D
 Acq On : 18 Sep 2023 19:08
 Operator : JC/SY
 Sample : 04436-01
 Misc : 5.97G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B103-01(19.5-20)

Quant Time: Sep 19 04:59:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	107464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	271183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	301741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	118096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	71966	66.356	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.720%
35) Dibromofluoromethane	7.810	113	99850	53.010	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.020%
50) Toluene-d8	10.269	98	352982	52.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.300%
62) 4-Bromofluorobenzene	12.575	95	104965	50.511	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.020%
Target Compounds						
						Qvalue
16) Acetone	4.022	43	6603	26.638	ug/l #	80
17) Carbon Disulfide	4.269	76	221590	53.845	ug/l	99
20) Methylene Chloride	4.799	84	21297	7.968	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091823\
Data File : VD077138.D
Acq On : 18 Sep 2023 19:08
Operator : JC/SY
Sample : 04436-01
Misc : 5.97G/5.0ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B103-01(19.5-20)

Quant Time: Sep 19 04:59:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

