


Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091918\
Data File : VD060035.D
Acq On : 19 Sep 2018 9:57
Operator : JC/SY
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091318S.M
Title : SW846 8260
Last Update : Thu Sep 13 16:17:40 2018



AutoFind: Scans 1107, 1108, 1109; Background Corrected with Scan 1103

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	99365	PASS
75	95	30	60	40.6	203610	PASS
95	95	100	100	100.0	501952	PASS
96	95	5	9	6.4	32135	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.9	446232	PASS
175	174	5	9	7.1	31833	PASS
176	174	95	101	96.7	431506	PASS
177	176	5	9	6.7	28854	PASS