

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091923\
 Data File : VD077147.D
 Acq On : 19 Sep 2023 15:36
 Operator : JC/SY
 Sample : 04445-01RE
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-09142023RE

Quant Time: Sep 19 16:26:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	61901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	161390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	203005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	101020	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	56342	90.189	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	180.380%#
35) Dibromofluoromethane	7.805	113	66160	59.019	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.040%
50) Toluene-d8	10.269	98	213258	53.447	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.900%
62) 4-Bromofluorobenzene	12.575	95	78410	63.402	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	126.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091923\
Data File : VD077147.D
Acq On : 19 Sep 2023 15:36
Operator : JC/SY
Sample : 04445-01RE
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-09142023RE

Quant Time: Sep 19 16:26:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

