

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091923\
 Data File : VD077149.D
 Acq On : 19 Sep 2023 16:32
 Operator : JC/SY
 Sample : 04433-01
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PILE-2

Quant Time: Sep 20 13:44:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	27413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	72096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	94855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	49013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	28263	102.160	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 204.320%#	
35) Dibromofluoromethane	7.805	113	31050	62.004	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.000%	
50) Toluene-d8	10.269	98	96043	53.882	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.760%	
62) 4-Bromofluorobenzene	12.575	95	35875	64.936	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 129.880%	
Target Compounds						Qvalue
20) Methylene Chloride	4.793	84	9696	19.906	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091923\
Data File : VD077149.D
Acq On : 19 Sep 2023 16:32
Operator : JC/SY
Sample : 04433-01
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PILE-2

Quant Time: Sep 20 13:44:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

