


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091923\
Data File : VD077142.D
Acq On : 19 Sep 2023 09:14
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Title : SW846 8260
Last Update : Tue Sep 19 04:46:37 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1858

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.9	24288	PASS
75	95	30	60	52.2	71013	PASS
95	95	100	100	100.0	136056	PASS
96	95	5	9	6.9	9452	PASS
173	174	0.00	2	1.4	1809	PASS
174	95	50	100	92.4	125675	PASS
175	174	5	9	6.9	8621	PASS
176	174	95	101	98.3	123536	PASS
177	176	5	9	6.9	8548	PASS