

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
 Data File : VD070382.D
 Acq On : 20 Sep 2021 14:20
 Operator : VA/SY
 Sample : M3792-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RT4213

Quant Time: Sep 21 03:33:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	125621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	259575	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	348828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	107865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	107086	66.389	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 132.780%	
35) Dibromofluoromethane	7.914	113	104530	56.573	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.140%	
50) Toluene-d8	10.338	98	357275	53.257	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.520%	
62) 4-Bromofluorobenzene	12.626	95	162079	69.131	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 138.260%	
Target Compounds						
16) Acetone	4.156	43	8475	29.429	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
Data File : VD070382.D
Acq On : 20 Sep 2021 14:20
Operator : VA/SY
Sample : M3792-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RT4213

Quant Time: Sep 21 03:33:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

