

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
 Data File : VD070385.D
 Acq On : 20 Sep 2021 15:45
 Operator : VA/SY
 Sample : M3798-04
 Misc : 7.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NAPL-07-(22-24)

Quant Time: Sep 21 03:34:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	200243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	403103	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	588087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	322151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	158044	61.468	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.940%	
35) Dibromofluoromethane	7.909	113	153629	53.542	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.080%	
50) Toluene-d8	10.338	98	559830	53.737	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.480%	
62) 4-Bromofluorobenzene	12.620	95	343876	94.448	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 188.900%#	
Target Compounds						
						Qvalue
16) Acetone	4.162	43	6384	13.907	ug/l	# 79
17) Carbon Disulfide	4.420	76	29260	3.402	ug/l	100
27) cis-1,2-Dichloroethene	7.209	96	27808	9.328	ug/l	89
44) Trichloroethene	9.108	130	14357	4.758	ug/l	92
64) Tetrachloroethene	10.873	164	86488	24.155	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
Data File : VD070385.D
Acq On : 20 Sep 2021 15:45
Operator : VA/SY
Sample : M3798-04
Misc : 7.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NAPL-07-(22-24)

Quant Time: Sep 21 03:34:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

