

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
 Data File : VD070375.D
 Acq On : 20 Sep 2021 10:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 03:32:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.574	0.470	18.1	90	0.00
3 P	Chloromethane	0.771	0.619	19.7	88	0.00
4 C	Vinyl Chloride	0.787	0.686	12.8#	91	0.00
5 T	Bromomethane	0.580	0.433	25.3#	87	0.00
6 T	Chloroethane	0.541	0.456	15.7	93	0.00
7 T	Trichlorofluoromethane	1.041	0.950	8.7	89	0.00
8 T	Diethyl Ether	0.288	0.253	12.2	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.646	0.590	8.7	88	0.00
10 T	Methyl Iodide	0.564	0.538	4.6	77	0.00
11 T	Tert butyl alcohol	0.073	0.026	64.4#	47#	0.00
12 CM	1,1-Dichloroethene	0.582	0.516	11.3#	83	0.00
13 T	Acrolein	0.032	0.024	25.0	71	0.00
14 T	Allyl chloride	0.855	0.740	13.5	83	0.00
15 T	Acrylonitrile	0.137	0.124	9.5	87	0.00
16 T	Acetone	0.115	0.117	-1.7	103	0.00
17 T	Carbon Disulfide	2.148	1.851	13.8	84	0.00
18 T	Methyl Acetate	0.344	0.295	14.2	91	0.00
19 T	Methyl tert-butyl Ether	1.258	1.165	7.4	82	0.00
20 T	Methylene Chloride	0.913	0.703	23.0	87	0.00
21 T	trans-1,2-Dichloroethene	0.693	0.637	8.1	85	0.00
22 T	Diisopropyl ether	1.831	1.761	3.8	89	0.00
23 T	Vinyl Acetate	0.848	0.820	3.3	89	0.00
24 P	1,1-Dichloroethane	1.295	1.165	10.0	88	0.00
25 T	2-Butanone	0.153	0.146	4.6	95	0.00
26 T	2,2-Dichloropropane	1.047	0.962	8.1	91	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.687	7.7	85	0.00
28 T	Bromochloromethane	0.514	0.484	5.8	92	0.00
29 T	Tetrahydrofuran	0.098	0.094	4.1	92	0.00
30 C	Chloroform	1.328	1.206	9.2#	89	0.00
31 T	Cyclohexane	1.106	1.006	9.0	88	0.00
32 T	1,1,1-Trichloroethane	1.118	1.026	8.2	89	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.581	9.5	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.356	0.344	3.4	88	0.00
36 T	1,1-Dichloropropene	0.534	0.532	0.4	88	0.00
37 T	Ethyl Acetate	0.200	0.199	0.5	95	0.00
38 T	Carbon Tetrachloride	0.529	0.534	-0.9	90	0.00
39 T	Methylcyclohexane	0.597	0.620	-3.9	86	0.00
40 TM	Benzene	1.566	1.535	2.0	86	0.00
41 T	Methacrylonitrile	0.105	0.093	11.4	85	0.00
42 TM	1,2-Dichloroethane	0.419	0.410	2.1	91	0.00
43 T	Isopropyl Acetate	0.370	0.343	7.3	87	0.00
44 TM	Trichloroethene	0.374	0.364	2.7	84	0.00
45 C	1,2-Dichloropropane	0.411	0.408	0.7#	90	0.00
46 T	Dibromomethane	0.212	0.209	1.4	90	0.00
47 T	Bromodichloromethane	0.540	0.532	1.5	89	0.00
48 T	Methyl methacrylate	0.171	0.173	-1.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
 Data File : VD070375.D
 Acq On : 20 Sep 2021 10:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 03:32:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	83	0.00
50 S	Toluene-d8	1.292	1.297	-0.4	86	0.00
51 T	4-Methyl-2-Pentanone	0.191	0.196	-2.6	92	0.00
52 CM	Toluene	0.946	0.977	-3.3#	86	0.00
53 T	t-1,3-Dichloropropene	0.464	0.477	-2.8	90	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.567	-0.4	90	0.00
55 T	1,1,2-Trichloroethane	0.291	0.292	-0.3	88	0.00
56 T	Ethyl methacrylate	0.318	0.326	-2.5	84	0.00
57 T	1,3-Dichloropropane	0.496	0.495	0.2	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.159	-3.9	90	0.00
59 T	2-Hexanone	0.128	0.139	-8.6	96	0.00
60 T	Dibromochloromethane	0.339	0.339	0.0	86	0.00
61 T	1,2-Dibromoethane	0.268	0.264	1.5	87	0.00
62 S	4-Bromofluorobenzene	0.452	0.436	3.5	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.304	0.308	-1.3	86	0.00
65 PM	Chlorobenzene	1.036	1.016	1.9	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.388	-2.4	89	0.00
67 C	Ethyl Benzene	1.786	1.869	-4.6#	86	0.00
68 T	m/p-Xylenes	0.690	0.740	-7.2	86	0.00
69 T	o-Xylene	0.618	0.662	-7.1	85	0.00
70 T	Styrene	1.080	1.199	-11.0	86	0.00
71 P	Bromoform	0.187	0.186	0.5	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.506	3.568	-1.8	87	0.00
74 T	N-amyl acetate	0.764	0.729	4.6	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.751	0.690	8.1	91	0.00
76 T	1,2,3-Trichloropropane	0.506	0.454	10.3	86	0.00
77 T	Bromobenzene	0.809	0.766	5.3	86	0.00
78 T	n-propylbenzene	4.551	4.656	-2.3	90	0.00
79 T	2-Chlorotoluene	2.653	2.621	1.2	88	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.120	-3.1	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.206	-2.0	96	0.00
82 T	4-Chlorotoluene	2.799	2.779	0.7	88	0.00
83 T	tert-Butylbenzene	2.427	2.492	-2.7	88	0.00
84 T	1,2,4-Trimethylbenzene	2.980	3.093	-3.8	88	0.00
85 T	sec-Butylbenzene	3.858	3.950	-2.4	89	0.00
86 T	p-Isopropyltoluene	3.100	3.236	-4.4	88	0.00
87 T	1,3-Dichlorobenzene	1.684	1.616	4.0	87	0.00
88 T	1,4-Dichlorobenzene	1.691	1.614	4.6	89	0.00
89 T	n-Butylbenzene	3.091	3.134	-1.4	89	0.00
90 T	Hexachloroethane	0.710	0.659	7.2	91	0.00
91 T	1,2-Dichlorobenzene	1.486	1.409	5.2	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.098	12.5	87	0.00
93 T	1,2,4-Trichlorobenzene	0.802	0.754	6.0	86	0.00
94 T	Hexachlorobutadiene	0.492	0.473	3.9	95	0.00
95 T	Naphthalene	1.458	1.402	3.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092021\
Data File : VD070375.D
Acq On : 20 Sep 2021 10:52
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 21 03:32:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.708	0.670	5.4	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6