

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092023\
 Data File : VD077161.D
 Acq On : 20 Sep 2023 10:59
 Operator : JC/SY
 Sample : VD0920SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0920SBL01

Quant Time: Sep 21 02:21:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	212284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	510576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	603882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	297746	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	119169	55.624	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.240%
35) Dibromofluoromethane	7.805	113	173491	48.920	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.840%
50) Toluene-d8	10.269	98	679355	53.818	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.640%
62) 4-Bromofluorobenzene	12.575	95	243536	62.246	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	124.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092023\
Data File : VD077161.D
Acq On : 20 Sep 2023 10:59
Operator : JC/SY
Sample : VD0920SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0920SBL01

Quant Time: Sep 21 02:21:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

