

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092023\
 Data File : VD077166.D
 Acq On : 20 Sep 2023 14:52
 Operator : JC/SY
 Sample : 04479-06
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 1028

Quant Time: Sep 21 02:24:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	227438	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	507487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	601979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	288233	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	116260	50.651	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.300%
35) Dibromofluoromethane	7.804	113	164456	46.655	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.300%
50) Toluene-d8	10.269	98	663860	52.911	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.820%
62) 4-Bromofluorobenzene	12.575	95	237540	61.083	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	122.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092023\
Data File : VD077166.D
Acq On : 20 Sep 2023 14:52
Operator : JC/SY
Sample : 04479-06
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
1028

Quant Time: Sep 21 02:24:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

