

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
 Data File : VD070397.D
 Acq On : 21 Sep 2021 11:49
 Operator : VA/SY
 Sample : VD0921SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0921SBL01

Quant Time: Sep 22 01:46:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	411959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	798284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	740103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	309530	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	261864	49.505	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.020%
35) Dibromofluoromethane	7.908	113	266746	46.943	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.880%
50) Toluene-d8	10.338	98	919351	44.562	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.120%
62) 4-Bromofluorobenzene	12.626	95	297060	41.200	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
Data File : VD070397.D
Acq On : 21 Sep 2021 11:49
Operator : VA/SY
Sample : VD0921SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0921SBL01

Quant Time: Sep 22 01:46:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

