

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
 Data File : VD070406.D
 Acq On : 21 Sep 2021 16:07
 Operator : VA/SY
 Sample : M3823-03
 Misc : 10.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DKM-ENV-IG-GRAB

Quant Time: Sep 22 01:48:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	268085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	527736	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	485408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	186748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	205559	59.716	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.440%
35) Dibromofluoromethane	7.914	113	190270	50.651	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.300%
50) Toluene-d8	10.338	98	525040	38.496	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.000%
62) 4-Bromofluorobenzene	12.620	95	182565	38.301	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.600%
Target Compounds						
20) Methylene Chloride	4.962	84	70537	13.051	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
Data File : VD070406.D
Acq On : 21 Sep 2021 16:07
Operator : VA/SY
Sample : M3823-03
Misc : 10.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DKM-ENV-IG-GRAB

Quant Time: Sep 22 01:48:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

