

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
 Data File : VD070411.D
 Acq On : 21 Sep 2021 18:28
 Operator : VA/SY
 Sample : M3860-01
 Misc : 6.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PT-SOIL

Quant Time: Sep 22 01:48:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	235678	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	467558	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	408721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	124801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	156521	51.723	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.440%
35) Dibromofluoromethane	7.914	113	157477	47.317	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.640%
50) Toluene-d8	10.338	98	444466	36.782	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.560%
62) 4-Bromofluorobenzene	12.626	95	119868	28.384	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	56.760%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.961	84	48974	8.520	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
Data File : VD070411.D
Acq On : 21 Sep 2021 18:28
Operator : VA/SY
Sample : M3860-01
Misc : 6.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PT-SOIL

Quant Time: Sep 22 01:48:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

