



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092121\
Data File : VD070395.D
Acq On : 21 Sep 2021 09:37
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Title : SW846 8260
Last Update : Thu Sep 09 03:07:09 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1799

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.0	47547	PASS
75	95	30	60	48.7	144592	PASS
95	95	100	100	100.0	296992	PASS
96	95	5	9	6.6	19617	PASS
173	174	0.00	2	1.3	2693	PASS
174	95	50	100	71.7	213013	PASS
175	174	5	9	7.6	16226	PASS
176	174	95	101	96.3	205141	PASS
177	176	5	9	7.2	14714	PASS