

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092123\
 Data File : VD077174.D
 Acq On : 21 Sep 2023 12:05
 Operator : JC/SY
 Sample : 04527-01
 Misc : 5.13G/10.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EZYG8

Quant Time: Sep 22 01:30:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM090123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 22 01:29:35 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	211696	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	170512	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	53847	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.276	65	148061	28.384	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 113.520%		
7) Chloroethane-d5	2.805	69	135861	31.672	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 126.680%		
11) 1,1-Dichloroethene-d2	3.917	65	32947	27.805	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 111.240%#		
21) 2-Butanone-d5	6.987	46	49615	75.501	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 151.000%#		
24) Chloroform-d	7.563	84	177330	30.613	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 122.440%		
26) 1,2-Dichloroethane-d4	8.228	65	95830	34.904	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 139.600%#		
32) Benzene-d6	8.199	84	338409	35.934	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 143.720%#		
36) 1,2-Dichloropropane-d6	9.210	67	108745	40.264	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 161.040%#		
41) Toluene-d8	10.269	98	248329	28.806	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 115.240%		
43) trans-1,3-Dichloroprop...	10.522	79	39935	32.453	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 129.800%		
47) 2-Hexanone-d5	10.875	63	35039	81.231	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 162.460%#		
56) 1,1,2,2-Tetrachloroeth...	12.651	84	86162	35.786	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 143.160%#		
66) 1,2-Dichlorobenzene-d4	13.816	152	44565	25.785	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 103.120%		
Target Compounds						
						Qvalue
14) Carbon disulfide	4.264	76	28836	3.134	ug/L	99
16) Methylene chloride	4.793	84	24724	5.738	ug/L	98
46) Tetrachloroethene	10.804	164	4066	1.831	ug/L	79
52) Ethylbenzene	11.681	91	37106	3.090	ug/L	95
53) m,p-Xylene	11.793	106	73154	15.779	ug/L	97
54) o-Xylene	12.116	106	30262	6.762	ug/L	98
63) 1,2,4-Trimethylbenzene	13.216	105	5106	0.971	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092123\
Data File : VD077174.D
Acq On : 21 Sep 2023 12:05
Operator : JC/SY
Sample : 04527-01
Misc : 5.13G/10.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EZYG8

Quant Time: Sep 22 01:30:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM090123SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 22 01:29:35 2023
Response via : Initial Calibration

