

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD092220\
 Data File : VD066815.D
 Acq On : 22 Sep 2020 17:13
 Operator : VA/SY
 Sample : L3872-07
 Misc : 5.21G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BU-02-092120

Quant Time: Sep 23 02:49:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	396430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	737917	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	719954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	309662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	239333	61.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.36%	
35) Dibromofluoromethane	7.91	113	246043	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
50) Toluene-d8	10.34	98	908149	56.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.56%	
62) 4-Bromofluorobenzene	12.63	95	297052	54.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.50%	
Target Compounds						
16) Acetone	4.15	43	21291	25.867	ug/l	91
20) Methylene Chloride	4.97	84	51290	6.479	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD092220\
Data File : VD066815.D
Acq On : 22 Sep 2020 17:13
Operator : VA/SY
Sample : L3872-07
Misc : 5.21G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BU-02-092120

Quant Time: Sep 23 02:49:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

