

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092220\
 Data File : VD066808.D
 Acq On : 22 Sep 2020 13:59
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD092220

Quant Time: Sep 23 02:21:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.412	0.362	12.1	105	0.00
3 P	Chloromethane	0.459	0.448	2.4	107	0.00
4 C	Vinyl Chloride	0.454	0.450	0.9#	105	0.00
5 T	Bromomethane	0.319	0.344	-7.8	113	0.00
6 T	Chloroethane	0.282	0.278	1.4	104	0.00
7 T	Trichlorofluoromethane	0.932	0.901	3.3	102	0.00
8 T	Diethyl Ether	0.237	0.237	0.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.544	-0.7	107	-0.01
10 T	Methyl Iodide	0.477	0.501	-5.0	101	0.00
11 T	Tert butyl alcohol	0.040	0.031	22.5#	98	0.00
12 CM	1,1-Dichloroethene	0.490	0.493	-0.6#	100	0.00
13 T	Acrolein	0.036	0.034	5.6	95	0.00
14 T	Allyl chloride	0.719	0.762	-6.0	103	0.00
15 T	Acrylonitrile	0.105	0.108	-2.9	98	0.00
16 T	Acetone	0.104	0.104	0.0	94	0.00
17 T	Carbon Disulfide	1.497	1.504	-0.5	101	0.00
18 T	Methyl Acetate	0.258	0.236	8.5	96	0.00
19 T	Methyl tert-butyl Ether	1.063	1.117	-5.1	99	0.00
20 T	Methylene Chloride	0.675	0.607	10.1	103	0.00
21 T	trans-1,2-Dichloroethene	0.567	0.580	-2.3	102	0.00
22 T	Diisopropyl ether	1.478	1.597	-8.1	102	0.00
23 T	Vinyl Acetate	0.727	0.856	-17.7	100	0.00
24 P	1,1-Dichloroethane	0.989	1.005	-1.6	101	0.00
25 T	2-Butanone	0.125	0.126	-0.8	96	0.00
26 T	2,2-Dichloropropane	0.906	0.904	0.2	103	0.00
27 T	cis-1,2-Dichloroethene	0.616	0.640	-3.9	100	0.00
28 T	Bromochloromethane	0.409	0.454	-11.0	106	0.00
29 T	Tetrahydrofuran	0.080	0.083	-3.8	97	0.00
30 C	Chloroform	1.066	1.064	0.2#	101	0.00
31 T	Cyclohexane	0.890	0.874	1.8	103	0.00
32 T	1,1,1-Trichloroethane	0.964	0.986	-2.3	102	0.00
33 S	1,2-Dichloroethane-d4	0.489	0.491	-0.4	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.318	0.329	-3.5	103	0.00
36 T	1,1-Dichloropropene	0.496	0.513	-3.4	102	0.00
37 T	Ethyl Acetate	0.179	0.193	-7.8	98	0.00
38 T	Carbon Tetrachloride	0.542	0.552	-1.8	103	0.00
39 T	Methylcyclohexane	0.570	0.622	-9.1	105	0.00
40 TM	Benzene	1.382	1.431	-3.5	101	0.00
41 T	Methacrylonitrile	0.110	0.118	-7.3	103	0.00
42 TM	1,2-Dichloroethane	0.409	0.418	-2.2	102	0.00
43 T	Isopropyl Acetate	0.346	0.358	-3.5	98	0.00
44 TM	Trichloroethene	0.394	0.393	0.3	100	0.00
45 C	1,2-Dichloropropane	0.348	0.351	-0.9#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092220\
 Data File : VD066808.D
 Acq On : 22 Sep 2020 13:59
 Operator : VA/SY
 Sample : VSTDICV050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD092220

Quant Time: Sep 23 02:21:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.189	0.196	-3.7	98	0.00
47 T	Bromodichloromethane	0.508	0.518	-2.0	101	0.00
48 T	Methyl methacrylate	0.170	0.197	-15.9	110	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	93	0.00
50 S	Toluene-d8	1.084	1.172	-8.1	105	0.00
51 T	4-Methyl-2-Pentanone	0.175	0.187	-6.9	98	0.00
52 CM	Toluene	0.890	0.942	-5.8#	102	0.00
53 T	t-1,3-Dichloropropene	0.432	0.463	-7.2	100	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.551	-4.6	99	0.00
55 T	1,1,2-Trichloroethane	0.260	0.261	-0.4	100	0.00
56 T	Ethyl methacrylate	0.278	0.304	-9.4	98	0.00
57 T	1,3-Dichloropropane	0.429	0.448	-4.4	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.132	0.128	3.0	99	0.00
59 T	2-Hexanone	0.118	0.128	-8.5	97	0.00
60 T	Dibromochloromethane	0.339	0.342	-0.9	99	0.00
61 T	1,2-Dibromoethane	0.242	0.253	-4.5	101	0.00
62 S	4-Bromofluorobenzene	0.371	0.404	-8.9	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.360	0.361	-0.3	102	0.00
65 PM	Chlorobenzene	1.041	1.077	-3.5	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.405	-2.8	101	0.00
67 C	Ethyl Benzene	1.828	1.980	-8.3#	103	0.00
68 T	m/p-Xylenes	0.692	0.761	-10.0	105	0.00
69 T	o-Xylene	0.626	0.683	-9.1	103	0.00
70 T	Styrene	1.087	1.197	-10.1	103	0.00
71 P	Bromoform	0.199	0.206	-3.5	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.612	4.057	-12.3	106	0.00
74 T	N-amyl acetate	0.705	0.775	-9.9	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.623	0.637	-2.2	100	0.00
76 T	1,2,3-Trichloropropane	0.414	0.403	2.7	96	0.00
77 T	Bromobenzene	0.845	0.894	-5.8	102	0.00
78 T	n-propylbenzene	4.345	4.762	-9.6	104	0.00
79 T	2-Chlorotoluene	2.476	2.657	-7.3	106	0.00
80 T	1,3,5-Trimethylbenzene	3.065	3.396	-10.8	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.166	0.185	-11.4	101	0.00
82 T	4-Chlorotoluene	2.610	2.771	-6.2	103	0.00
83 T	tert-Butylbenzene	2.518	2.843	-12.9	108	0.00
84 T	1,2,4-Trimethylbenzene	3.039	3.372	-11.0	106	0.00
85 T	sec-Butylbenzene	3.622	3.978	-9.8	106	0.00
86 T	p-Isopropyltoluene	3.264	3.706	-13.5	108	0.00
87 T	1,3-Dichlorobenzene	1.669	1.768	-5.9	105	0.00
88 T	1,4-Dichlorobenzene	1.658	1.706	-2.9	104	0.00
89 T	n-Butylbenzene	3.103	3.460	-11.5	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092220\
Data File : VD066808.D
Acq On : 22 Sep 2020 13:59
Operator : VA/SY
Sample : VSTDICV050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ICVVD092220

Quant Time: Sep 23 02:21:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.663	0.694	-4.7	105	0.00
91 T	1,2-Dichlorobenzene	1.428	1.499	-5.0	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.102	-2.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.899	0.970	-7.9	105	0.00
94 T	Hexachlorobutadiene	0.539	0.582	-8.0	106	0.00
95 T	Naphthalene	1.526	1.717	-12.5	100	0.00
96 T	1,2,3-Trichlorobenzene	0.776	0.836	-7.7	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6