


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092220\
 Data File : VD066800.D
 Acq On : 22 Sep 2020 09:19
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Title : SW846 8260
 Last Update : Wed Sep 23 01:34:05 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	31957	PASS
75	95	30	60	53.5	90405	PASS
95	95	100	100	100.0	169066	PASS
96	95	5	9	6.5	10923	PASS
173	174	0.00	2	1.5	1920	PASS
174	95	50	100	76.1	128613	PASS
175	174	5	9	7.5	9601	PASS
176	174	95	101	97.6	125581	PASS
177	176	5	9	7.7	9638	PASS