

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092221\
 Data File : VD070420.D
 Acq On : 22 Sep 2021 12:08
 Operator : VA/SY
 Sample : M3824-01RE
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-BRE

Quant Time: Sep 23 05:30:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	8057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	16101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	17864	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	8532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	8778	84.850	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 169.700%#	
35) Dibromofluoromethane	7.914	113	4466	38.967	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 77.940%	
50) Toluene-d8	10.338	98	14141	33.983	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 67.960%	
62) 4-Bromofluorobenzene	12.626	95	5525	37.991	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 75.980%	
Target Compounds						Qvalue
16) Acetone	4.156	43	3134	169.679	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092221\
Data File : VD070420.D
Acq On : 22 Sep 2021 12:08
Operator : VA/SY
Sample : M3824-01RE
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-BRE

Quant Time: Sep 23 05:30:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

