

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092221\
 Data File : VD070421.D
 Acq On : 22 Sep 2021 12:36
 Operator : VA/SY
 Sample : M3823-03
 Misc : 10.84G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DKM-ENV-IG-GRAB

Quant Time: Sep 23 05:31:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	259928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	503993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	489856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	187473	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	190916	57.203	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.400%
35) Dibromofluoromethane	7.908	113	179024	49.902	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.800%
50) Toluene-d8	10.338	98	517654	39.742	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	79.480%
62) 4-Bromofluorobenzene	12.626	95	167015	36.689	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092221\
Data File : VD070421.D
Acq On : 22 Sep 2021 12:36
Operator : VA/SY
Sample : M3823-03
Misc : 10.84G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DKM-ENV-IG-GRAB

Quant Time: Sep 23 05:31:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

