

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092221\
 Data File : VD070424.D
 Acq On : 22 Sep 2021 14:01
 Operator : VA/SY
 Sample : M3870-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-C

Quant Time: Sep 23 05:31:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	527	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.844	114	831	50.000	ug/l	-0.02
63) Chlorobenzene-d5	11.638	117	516	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.57
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	253	37.388	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	74.780%
35) Dibromofluoromethane	7.914	113	328	55.451	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.900%
50) Toluene-d8	10.338	98	1180	54.944	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.880%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092221\
Data File : VD070424.D
Acq On : 22 Sep 2021 14:01
Operator : VA/SY
Sample : M3870-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-C

Quant Time: Sep 23 05:31:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

