

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD092418\
 Data File : VD060084.D
 Acq On : 24 Sep 2018 16:12
 Operator : VA/AP
 Sample : J5017-01RE
 Misc : 5.05µl/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RA-091918-N-NW-055RE

Quant Time: Sep 25 01:41:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091318S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 13 16:17:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.37	168	1609023	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.41	114	2315385	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.26	117	1623091	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.36	152	606933	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.75	65	409236	32.61	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	65.22%	
35) Dibromofluoromethane	6.29	113	557819	30.84	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	61.68%	
50) Toluene-d8	9.42	98	1314853	26.23	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	52.46%	
62) 4-Bromofluorobenzene	12.41	95	431655	22.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	45.08%	
Target Compounds						
16) Acetone	3.13	43	129933	53.634	ug/l	97
25) 2-Butanone	5.59	43	57746	8.833	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD092418\
Data File : VD060084.D
Acq On : 24 Sep 2018 16:12
Operator : VA/AP
Sample : J5017-01RE
Misc : 5.05a/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RA-091918-N-NW-055RE

Quant Time: Sep 25 01:41:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091318S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 13 16:17:40 2018
Response via : Initial Calibration

