

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
 Data File : VD070450.D
 Acq On : 24 Sep 2021 13:04
 Operator : VA/SY
 Sample : M3907-03
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-206

Quant Time: Sep 25 03:08:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	227132	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	448535	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	391318	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	117016	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	165346	56.69	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.38%
35) Dibromofluoromethane	7.914	113	155867	48.82	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.64%
50) Toluene-d8	10.338	98	449846	38.81	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.62%
62) 4-Bromofluorobenzene	12.626	95	117323	28.96	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	57.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
Data File : VD070450.D
Acq On : 24 Sep 2021 13:04
Operator : VA/SY
Sample : M3907-03
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-206

Quant Time: Sep 25 03:08:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

