

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
 Data File : VD070458.D
 Acq On : 24 Sep 2021 16:48
 Operator : VA/SY
 Sample : M3907-03RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-206RE

Quant Time: Sep 25 03:09:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	214747	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	410731	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	378629	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	122864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	165477	60.01	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 120.02%		
35) Dibromofluoromethane	7.914	113	158512	54.22	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 108.44%		
50) Toluene-d8	10.338	98	449492	42.34	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 84.68%		
62) 4-Bromofluorobenzene	12.626	95	122262	32.96	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 65.92%		
Target Compounds						Qvalue
20) Methylene Chloride	4.962	84	39472	6.55	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
Data File : VD070458.D
Acq On : 24 Sep 2021 16:48
Operator : VA/SY
Sample : M3907-03RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-206RE

Quant Time: Sep 25 03:09:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

