

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
 Data File : VD070461.D
 Acq On : 24 Sep 2021 18:13
 Operator : VA/SY
 Sample : M3920-04
 Misc : 6.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-839-R-SO-6-6.5-092321

Quant Time: Sep 25 03:09:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	130745	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	259205	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	277068	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	111052	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	126966	75.63	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.26%
35) Dibromofluoromethane	7.908	113	107853	58.46	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.92%
50) Toluene-d8	10.332	98	271095	40.47	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	80.94%
62) 4-Bromofluorobenzene	12.626	95	94864	40.52	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.04%
Target Compounds						
					Qvalue	
16) Acetone	4.167	43	8780	29.29	ug/l	94
95) Naphthalene	15.385	128	6517	2.01	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
Data File : VD070461.D
Acq On : 24 Sep 2021 18:13
Operator : VA/SY
Sample : M3920-04
Misc : 6.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-839-R-SO-6-6.5-092321

Quant Time: Sep 25 03:09:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

