

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
 Data File : VD070465.D
 Acq On : 24 Sep 2021 20:04
 Operator : VA/SY
 Sample : M3908-06
 Misc : 6.35G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SUB-SLAB-06-(6-2.6)

Quant Time: Oct 04 07:48:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	36405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	70735	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	79928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	32103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	51872	110.968	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 221.940%#	
35) Dibromofluoromethane	7.908	113	35472	70.451	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 140.900%	
50) Toluene-d8	10.338	98	78999	43.214	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 86.420%	
62) 4-Bromofluorobenzene	12.626	95	29686	46.465	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.920%	
Target Compounds						
16) Acetone	4.156	43	4642	55.622	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092421\
Data File : VD070465.D
Acq On : 24 Sep 2021 20:04
Operator : VA/SY
Sample : M3908-06
Misc : 6.35G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SUB-SLAB-06-(6-2.6)

Quant Time: Oct 04 07:48:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

