

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD092520\
 Data File : VD066875.D
 Acq On : 25 Sep 2020 15:44
 Operator : VA/SY
 Sample : L4132-03
 Misc : 7.20G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-20413

Manual Integrations
 APPROVED

MMDadoda
 9/29/2020 11:45:23 AM

Quant Time: Sep 26 04:28:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	410096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	721794	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	730528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	346773	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	212874	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
35) Dibromofluoromethane	7.91	113	228423	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
50) Toluene-d8	10.34	98	869693	55.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.18%	
62) 4-Bromofluorobenzene	12.63	95	323150	60.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.66%	
Target Compounds						
8) Diethyl Ether	3.70	74	3656m	1.884	ug/l	Qvalue
16) Acetone	4.16	43	8299	9.747	ug/l	92
20) Methylene Chloride	4.96	84	66237	9.724	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD092520\
Data File : VD066875.D
Acq On : 25 Sep 2020 15:44
Operator : VA/SY
Sample : L4132-03
Misc : 7.20G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-20413

Manual Integrations
APPROVED

MMDadoda
9/29/2020 11:45:23 AM

Quant Time: Sep 26 04:28:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

