

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD092520\
 Data File : VD066901.D
 Acq On : 26 Sep 2020 05:23
 Operator : VA/SY
 Sample : L4127-07
 Misc : 5.05G/5.00ml/MSVOA D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-2(11-13)

Quant Time: Sep 26 07:45:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	453367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	763598	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	762249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	369196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	218248	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
35) Dibromofluoromethane	7.91	113	238660	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
50) Toluene-d8	10.34	98	916003	55.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.68%	
62) 4-Bromofluorobenzene	12.63	95	371880	65.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.26%	
Target Compounds						
16) Acetone	4.16	43	23090	24.529	ug/l	89
17) Carbon Disulfide	4.41	76	16220	1.195	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD092520\
Data File : VD066901.D
Acq On : 26 Sep 2020 05:23
Operator : VA/SY
Sample : L4127-07
Misc : 5.05G/5.00ml/MSVOA D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-2(11-13)

Quant Time: Sep 26 07:45:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

