

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092520\
 Data File : VD066864.D
 Acq On : 25 Sep 2020 10:36
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 03:08:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.412	0.348	15.5	99	0.00
3 P	Chloromethane	0.459	0.402	12.4	95	0.00
4 C	Vinyl Chloride	0.454	0.428	5.7#	98	0.00
5 T	Bromomethane	0.319	0.306	4.1	99	0.00
6 T	Chloroethane	0.282	0.256	9.2	94	0.00
7 T	Trichlorofluoromethane	0.932	0.871	6.5	97	0.00
8 T	Diethyl Ether	0.237	0.225	5.1	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.514	4.8	99	0.00
10 T	Methyl Iodide	0.477	0.434	9.0	86	0.00
11 T	Tert butyl alcohol	0.040	0.027	32.5#	84	0.00
12 CM	1,1-Dichloroethene	0.490	0.466	4.9#	93	0.00
13 T	Acrolein	0.036	0.041	-13.9	112	0.00
14 T	Allyl chloride	0.719	0.685	4.7	91	0.00
15 T	Acrylonitrile	0.105	0.101	3.8	90	0.00
16 T	Acetone	0.104	0.108	-3.8	95	0.00
17 T	Carbon Disulfide	1.497	1.380	7.8	91	0.00
18 T	Methyl Acetate	0.258	0.232	10.1	93	0.00
19 T	Methyl tert-butyl Ether	1.063	1.055	0.8	92	0.00
20 T	Methylene Chloride	0.675	0.554	17.9	92	0.00
21 T	trans-1,2-Dichloroethene	0.567	0.556	1.9	96	-0.01
22 T	Diisopropyl ether	1.478	1.472	0.4	92	0.00
23 T	Vinyl Acetate	0.727	0.774	-6.5	89	0.00
24 P	1,1-Dichloroethane	0.989	0.921	6.9	91	0.00
25 T	2-Butanone	0.125	0.120	4.0	90	0.00
26 T	2,2-Dichloropropane	0.906	0.842	7.1	94	0.00
27 T	cis-1,2-Dichloroethene	0.616	0.605	1.8	93	0.00
28 T	Bromochloromethane	0.409	0.407	0.5	93	0.00
29 T	Tetrahydrofuran	0.080	0.081	-1.3	94	0.00
30 C	Chloroform	1.066	1.008	5.4#	94	0.00
31 T	Cyclohexane	0.890	0.816	8.3	94	0.00
32 T	1,1,1-Trichloroethane	0.964	0.952	1.2	97	0.00
33 S	1,2-Dichloroethane-d4	0.489	0.464	5.1	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.318	0.323	-1.6	95	0.00
36 T	1,1-Dichloropropene	0.496	0.514	-3.6	96	0.00
37 T	Ethyl Acetate	0.179	0.192	-7.3	91	0.00
38 T	Carbon Tetrachloride	0.542	0.565	-4.2	98	0.00
39 T	Methylcyclohexane	0.570	0.611	-7.2	96	0.00
40 TM	Benzene	1.382	1.431	-3.5	94	0.00
41 T	Methacrylonitrile	0.110	0.116	-5.5	95	0.00
42 TM	1,2-Dichloroethane	0.409	0.414	-1.2	94	0.00
43 T	Isopropyl Acetate	0.346	0.351	-1.4	90	0.00
44 TM	Trichloroethene	0.394	0.405	-2.8	96	0.00
45 C	1,2-Dichloropropane	0.348	0.356	-2.3#	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092520\
 Data File : VD066864.D
 Acq On : 25 Sep 2020 10:36
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 03:08:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.189	0.198	-4.8	93	0.00
47 T	Bromodichloromethane	0.508	0.529	-4.1	96	0.00
48 T	Methyl methacrylate	0.170	0.190	-11.8	99	0.00
49 T	1,4-Dioxane	0.002	0.003	-50.0#	104	0.00
50 S	Toluene-d8	1.084	1.156	-6.6	96	0.00
51 T	4-Methyl-2-Pentanone	0.175	0.189	-8.0	93	0.00
52 CM	Toluene	0.890	0.950	-6.7#	96	0.00
53 T	t-1,3-Dichloropropene	0.432	0.463	-7.2	93	0.00
54 T	cis-1,3-Dichloropropene	0.527	0.563	-6.8	95	0.00
55 T	1,1,2-Trichloroethane	0.260	0.272	-4.6	97	0.00
56 T	Ethyl methacrylate	0.278	0.306	-10.1	92	0.00
57 T	1,3-Dichloropropane	0.429	0.460	-7.2	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.132	0.128	3.0	92	0.00
59 T	2-Hexanone	0.118	0.134	-13.6	95	0.00
60 T	Dibromochloromethane	0.339	0.359	-5.9	97	0.00
61 T	1,2-Dibromoethane	0.242	0.260	-7.4	96	0.00
62 S	4-Bromofluorobenzene	0.371	0.401	-8.1	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.360	0.383	-6.4	104	0.00
65 PM	Chlorobenzene	1.041	1.072	-3.0	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.409	-3.8	97	0.00
67 C	Ethyl Benzene	1.828	1.963	-7.4#	98	0.00
68 T	m/p-Xylenes	0.692	0.764	-10.4	100	0.00
69 T	o-Xylene	0.626	0.673	-7.5	97	0.00
70 T	Styrene	1.087	1.197	-10.1	98	0.00
71 P	Bromoform	0.199	0.214	-7.5	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.612	3.886	-7.6	98	0.00
74 T	N-amyl acetate	0.705	0.734	-4.1	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.623	0.640	-2.7	97	0.00
76 T	1,2,3-Trichloropropane	0.414	0.489	-18.1	113	0.00
77 T	Bromobenzene	0.845	0.876	-3.7	97	0.00
78 T	n-propylbenzene	4.345	4.693	-8.0	100	0.00
79 T	2-Chlorotoluene	2.476	2.576	-4.0	99	0.00
80 T	1,3,5-Trimethylbenzene	3.065	3.330	-8.6	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.166	0.179	-7.8	95	0.00
82 T	4-Chlorotoluene	2.610	2.744	-5.1	99	0.00
83 T	tert-Butylbenzene	2.518	2.750	-9.2	101	0.00
84 T	1,2,4-Trimethylbenzene	3.039	3.307	-8.8	101	0.00
85 T	sec-Butylbenzene	3.622	3.944	-8.9	102	0.00
86 T	p-Isopropyltoluene	3.264	3.666	-12.3	103	0.00
87 T	1,3-Dichlorobenzene	1.669	1.749	-4.8	101	0.00
88 T	1,4-Dichlorobenzene	1.658	1.703	-2.7	101	0.00
89 T	n-Butylbenzene	3.103	3.385	-9.1	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092520\
Data File : VD066864.D
Acq On : 25 Sep 2020 10:36
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 26 03:08:38 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.663	0.682	-2.9	100	0.00
91 T	1,2-Dichlorobenzene	1.428	1.489	-4.3	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.101	-1.0	96	0.00
93 T	1,2,4-Trichlorobenzene	0.899	0.966	-7.5	101	0.00
94 T	Hexachlorobutadiene	0.539	0.577	-7.1	102	0.00
95 T	Naphthalene	1.526	1.697	-11.2	96	0.00
96 T	1,2,3-Trichlorobenzene	0.776	0.843	-8.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6