

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD092820\
 Data File : VD066908.D
 Acq On : 28 Sep 2020 12:52
 Operator : VA/SY
 Sample : L4130-01
 Misc : 7.97G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D-84

Quant Time: Sep 29 01:14:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 23 01:34:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	373796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	637422	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	562898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	164609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	188725	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
35) Dibromofluoromethane	7.91	113	202717	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
50) Toluene-d8	10.34	98	708981	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	12.63	95	205876	43.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.06%	
Target Compounds						
17) Carbon Disulfide	4.41	76	28190	2.519	ug/l	97
20) Methylene Chloride	4.97	84	61007	9.896	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD092820\
Data File : VD066908.D
Acq On : 28 Sep 2020 12:52
Operator : VA/SY
Sample : L4130-01
Misc : 7.97G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D-84

Quant Time: Sep 29 01:14:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D092220S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 23 01:34:05 2020
Response via : Initial Calibration

