

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
 Data File : VD070495.D
 Acq On : 28 Sep 2021 11:45
 Operator : VA/SY
 Sample : M3943-04
 Misc : 6.98G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-TV-1

Quant Time: Sep 28 23:40:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	100622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	206013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	250258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	97300	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	94423	73.082	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.160%
35) Dibromofluoromethane	7.914	113	91710	62.540	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.080%
50) Toluene-d8	10.338	98	274509	51.558	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.626	95	105445	56.668	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
Data File : VD070495.D
Acq On : 28 Sep 2021 11:45
Operator : VA/SY
Sample : M3943-04
Misc : 6.98G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-TV-1

Quant Time: Sep 28 23:40:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

