

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
 Data File : VD070497.D
 Acq On : 28 Sep 2021 12:41
 Operator : VA/SY
 Sample : M3946-01RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-05-092421RE

Quant Time: Sep 28 23:41:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	7115	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	14957	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	15870	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	6950	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	6682	73.141	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 146.280%	
35) Dibromofluoromethane	7.920	113	5061	47.536	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 95.080%	
50) Toluene-d8	10.332	98	16907	43.738	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 87.480%	
62) 4-Bromofluorobenzene	12.626	95	6304	46.664	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 93.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
Data File : VD070497.D
Acq On : 28 Sep 2021 12:41
Operator : VA/SY
Sample : M3946-01RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-05-092421RE

Quant Time: Sep 28 23:41:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

