

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
 Data File : VD070498.D
 Acq On : 28 Sep 2021 13:09
 Operator : VA/SY
 Sample : M3941-01
 Misc : 4.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124019

Quant Time: Sep 28 23:41:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	270976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	587453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	673145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	308878	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	188567	54.195	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.400%
35) Dibromofluoromethane	7.908	113	215285	51.484	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.960%
50) Toluene-d8	10.332	98	785541	51.741	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.626	95	302172	56.949	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
Data File : VD070498.D
Acq On : 28 Sep 2021 13:09
Operator : VA/SY
Sample : M3941-01
Misc : 4.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124019

Quant Time: Sep 28 23:41:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

