

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
 Data File : VD070501.D
 Acq On : 28 Sep 2021 14:34
 Operator : VA/SY
 Sample : M3922-04
 Misc : 5.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SUB-SLAB-02-(43-45)

Quant Time: Sep 28 23:42:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	305917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	674346	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	740482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	328543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	228253	58.109	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.220%
35) Dibromofluoromethane	7.914	113	245105	51.063	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.120%
50) Toluene-d8	10.338	98	908461	52.127	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.260%
62) 4-Bromofluorobenzene	12.626	95	337867	55.471	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.940%
Target Compounds						
						Qvalue
4) Vinyl Chloride	2.350	62	15813	3.285	ug/l	# 81
16) Acetone	4.161	43	11875	16.933	ug/l	95
27) cis-1,2-Dichloroethene	7.202	96	6751	1.482	ug/l	88
64) Tetrachloroethene	10.867	164	59924	13.292	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
Data File : VD070501.D
Acq On : 28 Sep 2021 14:34
Operator : VA/SY
Sample : M3922-04
Misc : 5.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SUB-SLAB-02-(43-45)

Quant Time: Sep 28 23:42:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

