

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
 Data File : VD070511.D
 Acq On : 28 Sep 2021 19:17
 Operator : VA/SY
 Sample : M3938-05
 Misc : 6.75G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SUB-SLAB-08(6-8)

Quant Time: Sep 28 23:45:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	283686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	597629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	715085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	784049	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	217584	59.733	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.460%	
35) Dibromofluoromethane	7.914	113	228892	53.806	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.620%	
50) Toluene-d8	10.332	98	881230	57.055	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 114.120%	
62) 4-Bromofluorobenzene	12.620	95	1744413	323.165	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 646.320%#	
Target Compounds						Qvalue
16) Acetone	4.156	43	15432	23.729	ug/l	# 85
17) Carbon Disulfide	4.409	76	27076	2.222	ug/l	96
27) cis-1,2-Dichloroethene	7.209	96	11092	2.626	ug/l	89
44) Trichloroethene	9.108	130	15381	3.438	ug/l	83
64) Tetrachloroethene	10.867	164	67722	15.555	ug/l	97
85) sec-Butylbenzene	13.391	105	1102326	18.221	ug/l	# 51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092821\
Data File : VD070511.D
Acq On : 28 Sep 2021 19:17
Operator : VA/SY
Sample : M3938-05
Misc : 6.75G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SUB-SLAB-08(6-8)

Quant Time: Sep 28 23:45:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

