

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092823\
 Data File : VD077234.D
 Acq On : 28 Sep 2023 17:23
 Operator : JC/SY
 Sample : 04668-07
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR200027

Quant Time: Sep 29 02:20:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	21376	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	45870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	69550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	35880	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	15486	71.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.560%
35) Dibromofluoromethane	7.805	113	16024	50.293	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.580%
50) Toluene-d8	10.269	98	63554	56.041	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.080%
62) 4-Bromofluorobenzene	12.575	95	34584	98.391	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	196.780%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092823\
Data File : VD077234.D
Acq On : 28 Sep 2023 17:23
Operator : JC/SY
Sample : 04668-07
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR200027

Quant Time: Sep 29 02:20:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

