

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092823\
 Data File : VD077237.D
 Acq On : 28 Sep 2023 18:45
 Operator : JC/SY
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 29 02:21:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 19 04:46:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	283899	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	626631	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	816047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	440505	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	135364	47.245	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.500%
35) Dibromofluoromethane	7.804	113	191489	43.995	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.980%
50) Toluene-d8	10.269	98	836485	53.993	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.575	95	369355	76.920	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	153.840%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 29 02:21:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 19 04:46:37 2023
Response via : Initial Calibration

