



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092823\
Data File : VD077228.D
Acq On : 28 Sep 2023 08:58
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091823S.M
Title : SW846 8260
Last Update : Tue Sep 19 04:46:37 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	22509	PASS
75	95	30	60	50.4	65541	PASS
95	95	100	100	100.0	130117	PASS
96	95	5	9	6.9	8924	PASS
173	174	0.00	2	1.0	1211	PASS
174	95	50	100	92.4	120213	PASS
175	174	5	9	7.2	8704	PASS
176	174	95	101	95.0	114168	PASS
177	176	5	9	6.5	7437	PASS